

# **Validation of methods for prediction of clinical output levels of active middle ear implants from measurements in human cadaveric ears**

-

## **Supplementary Information**


Martin Grossöhmichen<sup>1,2,\*</sup>, Bernd Waldmann<sup>3</sup>, Rolf Salcher<sup>1,2</sup>, Nils Prenzler<sup>1,2</sup>, Thomas Lenarz<sup>1,2</sup>,  
Hannes Maier<sup>1,2</sup>

<sup>1</sup> Department of Otolaryngology and Institute of Audioneurotechnology (VIANNA), Hannover Medical School, 30625 Hannover, Germany


<sup>2</sup> DFG Cluster of Excellence “Hearing4all”, Germany

<sup>3</sup> Cochlear Deutschland GmbH & Co. KG, 30625 Hannover, Germany


\* Corresponding author: [grossoehmichen.martin@mh-hannover.de](mailto:grossoehmichen.martin@mh-hannover.de)



**Supplementary Figure 1:** Stapes displacement responses to sound stimulation at the tympanic membrane in temporal bone preparations used for experiments. The black dashed lines depict the limits given by Rosowski et al.<sup>3</sup>.



**Supplementary Figure 2:** Intracochlear pressure differences (ICPD) normalized to the outer ear canal sound pressure level ( $p_T$ ). For comparison the range of results<sup>16</sup>, obtained with a custom made pressure sensor are given as grey shaded area.



**Supplementary Figure 3:** Comparison of the T2 actuator output (eq. dB SPL<sub>FF</sub>) in TBs, calculated from stapes vibration amplitudes, between the present study, Devèze et al.<sup>4</sup> and Tringali et al.<sup>22</sup>. Output levels in Devèze et al. and Tringali et al. were given as eq. ear canal SPLs [eq. dB SPL<sub>TM</sub>] for nominally 1 V<sub>rms</sub> actuator input voltage and have been converted to eq. free field SPLs [eq. dB SPL<sub>FF</sub>] using tables I to III in Shaw et al.<sup>19</sup> to make them comparable to our data.